

Příloha 1 - Kalibrační protokol pro zkoušky tahem

Current Calibration Info

General

Calibration date **Tue May 10 14:45:47 2022**

Calibrated sensor

| | |
|------------------|-----------------------------|
| Sensor name | ARAMIS 4M (FG) |
| Measuring volume | Adjustable measuring volume |
| Camera support | Adjustable base 500 |
| Working distance | 780 mm |
| Camera angle | 25° |
| Camera distance | 318 mm |
| Serial number | no identifier |

Calibration object

| | |
|---------------------------|--|
| Object type | Panel (coded) |
| Name | GOM-CP20-MV55x44 |
| Calibration scale | Distance 1: 47.521 mm Distance 2: 47.526 mm |
| Certification temperature | 20.0 °C |
| Expansion coefficient | $4.00 \times 10^{-6} \text{ K}^{-1}$ |
| Calibration temperature | 20.0 °C |

Calibration settings

| | |
|-----------------------|-------------------------------|
| Camera lenses | 100.00 mm |
| Snap mode | Single snap |
| Max. ellipse residual | 0.088 (gray value adjustment) |

Calibration Result

| | |
|-------------------------------|--------------------------------|
| Calibration deviation | 0.027 Pixels |
| Calibration deviation (check) | OK (limit value: 0.050 Pixels) |
| Scale discrepancy | 0.003 mm |
| Scale discrepancy (check) | OK (limit value: 0.005 mm) |
| Camera angle | 24.8° |
| Height variance | 36 mm |
| Measuring volume | 100 / 75 / 55 mm |